

CERTIFICATE OF COMPLIANCE

Certificate Number 20150629-E201387
Report Reference E201387-A8-UL
Issue Date 2015-JUNE-29

Issued to: JUMO GMBH & CO KG
MORITZ-JUCHHEIM-STRASSE 1
36039 FULDA GERMANY

**This is to certify that
representative samples of**

COMPONENT - PROCESS CONTROL EQUIPMENT,
ELECTRICAL

Measuring, Control, and Automation System
JUMO mTRON T type 705010, type 705015, type 705060,
type 705001, type 705040, type 705020, type 705021, type
705025, type 705030

Have been investigated by UL in accordance with the
Standard(s) indicated on this Certificate.

Standard(s) for Safety: UL 61010-1 & CAN/CSA-C22.2 No. 61010-1 Standard For
Electrical Equipment For Measurement, Control, And
Laboratory Use

Additional Information: See the UL Online Certifications Directory at
www.ul.com/database for additional information

Only those products bearing the UL Certification Mark should be considered as being covered by UL's
Certification and Follow-Up Service.

Recognized components are incomplete in certain constructional features or restricted in performance
capabilities and are intended for use as components of complete equipment submitted for investigation rather
than for direct separate installation in the field. The final acceptance of the component is dependent upon its
installation and use in complete equipment submitted to UL LLC.

Look for the UL Certification Mark on the product.



Bruce Mahrenholz, Director North American Certification Program
UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please
contact a local UL Customer Service Representative at <http://ul.com/aboutul/locations/>

